

Bibliography

- [1] PETERS A., WICHMANN H.E., TUCH T., HEINRICH J., HEYDER J. Respiratory effects are associated with the number of ultrafine particles. *Am. J. Respir. Crit. Care Med.* 1997, **155** pp. 1376–1383
- [2] OBERDÖRSTER G. Toxicology of ultrafine particles: in vivo studies. *Philosophical Transactions of the Royal Society A*. 2000, **358** pp. 2719–2740
- [3] SCHMID O., STOEGER T. Surface area is the biologically most effective dose metric for acute nanoparticle toxicity in the lung. *J. Aerosol Sci.* 2016, **99** pp. 133–143
- [4] EN 12341:2023, *Ambient air — Standard gravimetric measurement method for the determination of the PM10 or PM2,5 mass concentration of suspended particulate matter*
- [5] EN 16976:2024, *Ambient air — Determination of the particle number concentration of atmospheric aerosol*
- [6] CEN/TS 17434:2020, *Ambient air — Determination of the particle number size distribution of atmospheric aerosol using a Mobility Particle Size Spectrometer (MPSS)*
- [7] ISO 27891:2015, *Aerosol particle number concentration — Calibration of condensation particle counters*
- [8] EN ISO 13138:2012, *Air quality — Sampling conventions for airborne particle deposition in the human respiratory system (ISO 13138:2012)*
- [9] ISO 16000-34:2018, *Indoor air — Part 34: Strategies for the measurement of airborne particles*
- [10] JUNG H., KITTELSON D.B. Characterization of Aerosol Surface Instruments in Transition Regime. *Aerosol Sci. Technol.* 2005, **39** pp. 902–911
- [11] FIERZ M., HOULE C., STEIGMEIER P., BURTSCHER H. Design, Calibration, and Field Performance of a Miniature Diffusion Size Classifier. *Aerosol Sci. Technol.* 2011, **45** pp. 1–10
- [12] KAMINSKI H., KUHLBUSCH T.A.J., FISSAN H., RAVI L., HORN H.G., HAN H.S. et al. Mathematical description of experimentally determined charge distributions of a unipolar diffusion charger. *Aerosol Sci. Technol.* 2012, **46** pp. 708–716
- [13] SHIN W.G., QI C., WANG J., FISSAN H., PUI D.Y.H. The effect of dielectric constant of materials on unipolar diffusion charging of nanoparticles. *J. Aerosol Sci.* 2009, **40** pp. 463–468
- [14] Klumpp, J.; Bertelli, L.: KDEP: A Resource for Calculating Particle Deposition in the Respiratory Tract, *Health Phys* Aug. 113(2) (2017), pp. 110–121
- [15] ICRP – International Commission on Radiological Protection Publication 66 *Human Respiratory Tract Model for Radiological Protection*. Pergamon, Elsevier Science Ltd, Oxford, 1994
- [16] FUCHS N.A. On the stationary charge distribution on aerosol particles in bipolar ionic atmosphere. *Geofis. Pura Appl.* 1963, **56** pp. 185–193
- [17] CHEN J., DAVIDSON J.H. Ozone production in the positive DC corona discharge: Model and comparison to experiments. *Plasma Chem. Plasma Process.* 2002, **22** pp. 495–522

- [18] BRANDVOLD D.K., MARTINEZ P., DOGRUEL D. Polarity dependence of N₂O formation from corona discharge. *Atmos. Environ.* 1989, **23** pp. 1881–1883
- [19] FIERZ M., MEIER D., STEIGMEIER P., BURTSCHER H. Aerosol Measurements by Induced Currents. *Aerosol Sci. Technol.* 2014, **48** pp. 350–357
- [20] <https://www.actris.eu/>
- [21] Directive 2008/50/EC of the European Parliament and of the Council of 21 May 2008 on ambient air quality and cleaner air for Europe
- [22] Swietlicki, E.; Hansson, H.-C.; Hämeri, K.; Svenningsson, B.; Massling, A.; McFiggans, G. et al.: Hygroscopic properties of submicrometer atmospheric aerosol particles measured with H-TDMA instruments in various environments – a review, Tellus B: Chemical and Physical Meteorology, 60:3 (2008), pp. 432-469, DOI: [10.1111/j.1600-0889.2008.00350.x](https://doi.org/10.1111/j.1600-0889.2008.00350.x)
- [23] LIU B.Y.H., PUI D.Y.H. A submicron aerosol standard and the primary, absolute calibration of the condensation nuclei counter. *J. Colloid Interface Sci.* 1974, **47** pp. 155–171
- [24] TAVAKOLI F., OLFERT J.S. An instrument for the classification of aerosols by particle relaxation time: Theoretical models of the Aerodynamic Aerosol Classifier. *Aerosol Sci. Technol.* 2013, **47** pp. 916–926
- [25] LÜÖND F., SCHLATTER J. Improved monodispersity of size-selected aerosol particles with a new charging selection scheme for tandem DMA setup. *J. Aerosol Sci.* 2013, **62** pp. 40–55
- [26] YOOK S.J., FISSAN H., ENGELKE T., ASBACH C., VAN DER ZWAAG T., KIM J.H. et al. Classification of highly monodisperse nanoparticles of NIST-traceable sizes by TDMA and control of deposition spot size on a surface by electrophoresis. *J. Aerosol Sci.* 2008, **39** pp. 537–548